

CIED implantation: Is it a just beginning of problems?

전남대학교병원 순환기내과
박형욱

Conflict of interest

Nothing to disclosure

Yes or No?

Multiple choices?

**Short-answer question
or essay question?**

Great job !!!



or



Oh My God !!!

The emerging use of **C**ardiac **I**mplantable **E**lectronic **D**evelopments (CIED)

- ✓ Resuscitation of the Heart in Ventricular Standstill by External Electric Stimulation.

Zoll PM. *N Engl J Med.* **1952**;247:768–771

- ✓ Elmqvist and Senning implanted the first cardiac pacemaker (PM) by thoracotomy.

Elmqvist R, Senning A. 24–27 June **1959**

- ✓ Termination of Malignant Ventricular Arrhythmias with an Implanted Automatic Defibrillator in Human Beings.

Mirowski M, et al. *N Engl J Med.* **1980**;303:322–324

Prevalence of infections in use of CEIDs - USA -

In the United States, from 1996 to 2003

Implantations; 0.49-fold

Infections; 3.1-fold


Year	CRT-D	CRT-P	AICD	Pacemaker	Other
1997	0 (0)	0 (0)	26,922 (2,333)	144,765 (5,416)	50,918 (2,735)
1998	0 (0)	0 (0)	28,260 (2,255)	147,695 (5,949)	41,387 (1,955)
1999	0 (0)	0 (0)	32,944 (3,168)	155,182 (6,438)	46,595 (2,711)
2000	0 (0)	0 (0)	39,334 (2,698)	164,845 (6,054)	49,488 (2,436)
2001	0 (0)	0 (0)	47,962 (3,988)	188,358 (7,096)	55,435 (3,092)
2002	1,623 (220)	734 (104)	66,528 (5,670)	188,224 (7,534)	67,593 (4,752)
2003	18,761 (1,714)	6,697 (536)	62,200 (4,190)	182,597 (6,590)	70,089 (2,643)
2004	32,737 (2,760)	7,325 (644)	66,545 (4,416)	178,816 (6,528)	60,118 (3,200)

Zhan C, et al. *J Gen Intern Med.* 2007;23(suppl 1):13–19

Between 2004 and 2006




Implantations; 12% increase

Infections; 57% increase



Uslan DZ, et al. *Arch Intern Med.* 2007;167:669–675

Prevalence of infections in use of CEIDs





- ✓ Infection rate; ICDs > PMs
- ✓ Hospitalizations; ICDs > PMs

- ✓ Longer lengths of stay
- ✓ Five-fold risk of in-hospital death
- ✓ High rates of one-year mortality even after the removal



Risk factors for CIEDs infections

- ✓ Immunosuppression (diabetes, CRF, steroid therapy and/or immunosuppressive therapy, malignancy)
- ✓ No periprocedural antimicrobial prophylaxis
- ✓ Fever less than 24 hours before implantation
- ✓ Vitamin K – antagonist therapy
- ✓ Postoperative hematoma
- ✓ Generator replacement
- ✓ Preprocedural temporary pacing
- ✓ Need for acute repositioning of a lead
- ✓ Presence of abandoned leads
- ✓ Operator experience
- ✓ Duration of procedure
- ✓ Number of implanted leads

32Y, M, ICD. Secondary prevention for DCMP VT



32Y, M, ICD. Secondary prevention for DCMP VT – Vegetation around lead



AHA Scientific Statement

Update on Cardiovascular Implantable Electronic Device Infections and Their Management

A Scientific Statement From the American Heart Association

Endorsed by the Heart Rhythm Society

- ✓ Once the diagnosis of CIED infection is made
- ✓ **Removal of both the device and the lead**
- ✓ High risk of relapse due to retained hardware, even in case of demonstrated valvular endocarditis without definite involvement of the leads or the device, and in case of persisting or relapsing SAB

CIEDs infections

- ✓ **Superficial or incisional pocket site infection**

Device involvement (-)

→ **No indication to remove** it, oral antibiotic therapy


Baddour LM, et al. *Circulation*. 2010;121:458–477.

- ✓ **Reimplantation**

; **No sooner than 72 hours after negative blood cultures**
Must be delayed by 2 weeks in case of demonstration of infectious involvement of a native valve


Tarakji K, et al. *Heart Rhythm*. 2010;7:1043–1047.

CIEDs infections



✓ Implanting the device with a polymer mesh

→ Releases minocycline and rifampin, in addition to standard systemic antimicrobial prophylaxis (cephalosporin or vancomycin).



- ✓ ***Conclusions:*** *CIED procedures that utilized an antibacterial envelope had a high rate of CIED implantation success (>99%). Although the follow-up to date is short, there was also a **low rate of infection (<0.50%) in this population at high risk for CIED infection.***

Safety is the first priority



Safety in the use of CEIDS

- ✓ CIEDs beget an intrinsic risk of **mechanical failure**
- ✓ A sharp increase in manufacturer **recalls**

Table 1. Current ICD Advisories Included in the Survey and Associated Risk

Company/Device*	Date of Advisory	Advisory Issue†	Current Risk of Failure, %‡
Medtronic Marquis ICD	February 2005	Accelerated battery depletion caused by internal battery short	0.01
Guidant Ventak Prizm 2 DR ICD	June 2005	Short circuit caused by wire insulation problem within lead connector block	0.1
Guidant Ventak Prizm AVT, Vitality AVT, and Contak Renewal AVT ICDs	June 2005	Random memory error, limiting delivery of therapies	0.0095
Guidant Contak Renewal 3, 4, Renewal 3, 4 AVT, and Renewal RF ICDs	June 2005	Magnetic switch faulty, impairing delivery of therapies	0.009
St Jude Photon DR, Photon Micro VR/DR, and Atlas VR/DR ICDs	October 2005	Memory chip affected by atmospheric radiation, which can impair pacing and delivery of therapies	0.167
ELA Alto ICD	August 2001	Migration of metal, which can impair pacing and delivery of therapies	2.6‡ 0.1§

Abbreviation: ICD, implantable cardioverter-defibrillator.

*Predominantly subpopulation of listed devices affected by advisory.

†Data obtained from physician communications and public statement releases such as those from Medtronic¹⁰ and Guidant.¹¹ The current risk of failure represents the number of failures divided by the number of devices implanted at the time of advisory disclosure.

‡Manufactured between April and July 2003.

§Manufactured between August 2003 and August 2004.

Safety in the use of CEIDS

Table 2. Advisory Device Replacement Population Characteristics (n = 533)


	Data
Total replacements, No.	533
Device manufacturer, %	
Medtronic	72
Guidant	27
St Jude	0.9
ELA	0.1
Patient age, mean (SD), y	64 (13)
Male, %	77
Secondary prevention indication for ICD, %	66
Previous appropriate shock, %	45
Pacing dependency, %	21

Table 3. Summary of Number of Advisory Devices and Number Replaced at Participating Implanting/Follow-up Centers

Site	No. of Advisory Devices	No. (%) Replaced	Devices Implanted/2 y*
1	229	14 (6)	317
2	138	62 (45)	407
3	378	61 (16)	439
4	131	30 (23)	428
5	248	21 (8)	359
6	153	24 (16)	312
7	90	33 (37)	297
8	130	47 (36)	693
9	59	23 (39)	170
10	94	22 (23)	213
11	15	2 (13)	22
12	410	90 (22)	699
13	196	32 (16)	380
14	63	0	149
15	53	24 (45)	0
16	177	21 (12)	449
17	351	27 (8)	405
Total	2915	533 (18.3)	5289

Safety in the use of CEIDS

- ✓ US FDA classifies medical device recalls; class I, II, III
- ✓ FDA and the HRS have set guidelines for advisories



U.S. Department of Health & Human Services

FDA U.S. Food and Drug Administration
Protecting and Promoting Your Health

Home Food Drugs Medical Devices Radiation-Emitting Products Vaccines, Blood & Biologics Animal & Veterinary Cosmetics Tobacco Products

Medical Devices

Medical Device Safety

List of Device Recalls

Search the Section

Medical Device Recalls: Get e-mail updates

FDA posts consumer information about the most serious medical device recalls. These products are on the list because there is a reasonable chance that they could cause serious health problems or death.

Use the yearly lists to find information about Class I medical device recalls and some Class II and III recalls of interest to consumers. The links give details about what to do if you own or use one of these products.

Please note that FDA now lists medical device recall notices by the date that it posts the recall rather than the recall initiation date. You can find the date that a firm initiated a recall in the text of the recall notice.

Learn more about medical device recalls here.

Recent Medical Device Recalls

Listed by date posted on FDA website.

Device Name	Date
Smith's Medical ASD, Inc., BCI Remote Alarm Cables (NO-Normally Open, NC-Normally Closed), used with 9004 the Capnocheck Capnography (Plus or Sleep) System, Reorder Numbers: WW3398NO & WW3398NC	04/09/13
Animas Corporation 2020 Insulin Infusion Pump	04/05/13
Vascular Solutions Inc., Guardian II and Guardian II NC Hemostasis Valves, Model Numbers 8210, 8211, 8215, and 8216	03/27/13
Spacelabs Medical, Inc., Spacelabs Anesthesia Workstations and Service Kits	03/26/13
Symbios Medical Products, LLC, GOPump Elastomeric Infusion Pump/KIT, Part Number 510076	03/18/13
Ad-Tech Medical Instrument Corporation, Macro Micro Subdural Electrodes	03/07/13
Lumenis Limited, Lumenis VersaCut Tissue Morcellator	02/25/13
DePuy Orthopaedics, Inc. - LPS Diaphyseal Sleeve	02/22/13

- ① Advisory has been issued
- ② Patients with an affected device must be contacted
- ③ Physicians have to choose the best management strategy
 - Estimates of rate of possible malfunction
 - Likely effect of the issue on specific patients (eg, pacemaker dependency)
 - Individual center's procedural risk associated with the replacement

Safety in the use of CEIDS

- ✓ Three management strategies are possible:

Replacing all the recalled devices

Replacing a few devices

Not replacing any device

- ✓ According to several analyses...
- ✓ Mean percentage of replaced devices, in different countries, is lower than 20%
- complications rate ranges from 0.62% to 8.1%

Safety in the use of CEIDS

- ✓ When a device is returned to the manufacturer;
Laboratory technicians and engineers assess all its functions and perform an **analysis** through a series of diagnostic tests that verify the performance of defibrillation, pacing, sensing, memory, and recording functions.
- ✓ **Test results are compared to original manufacturing records** and design intent. **Companies should inform the regulatory authorities** of each significant event that poses potential risk to patients' health, and **periodically publish a performance report**, indicating the overall incidence of malfunctions that have occurred for each product

Safety in the use of CEIDS

✓ July 2005,

✓ **FDA class I recall:**


Boston Scientific (formerly Guidant) about a potential short circuit that might have affected the **Prizm ICD** (model 1861), manufactured on or before 16 April 2002, and about a potential deterioration in the wire insulator within the lead connector block, affecting the **Renewal I** (model H135) and **Renewal II** (model H155) CRT-D, manufactured on or before 26 August 2004

The screenshot shows the FDA website interface. At the top, it says "U.S. Department of Health & Human Services" and "U.S. Food and Drug Administration". Below that is a navigation bar with links for Home, Food, Drugs, Medical Devices, Radiation-Emitting Products, Vaccines, Blood & Biologics, Animal & Veterinary, Cosmetics, and Tobacco Products. A search bar is also present. The main content area is titled "Medical Devices" and contains a section for "Medical Device Safety" with a sub-section for "Safety Communications". The primary heading is "Update of FDA Preliminary Public Health Notification*: Guidant VENTAK PRIZM® 2 DR and CONTAK RENEWAL® Implantable Cardioverter Defibrillators". Below this, it states "This is an archived document and is no longer current information." and "October 13, 2005". The text explains that this is an update to a July 14, 2005 PPHN about malfunctions with Guidant's PRIZM 2 and CONTAK RENEWAL implantable cardioverter defibrillator (ICD) devices. It lists the affected devices: VENTAK PRIZM 2 DR, Model 1861; CONTAK RENEWAL, Model H135; and CONTAK RENEWAL 2, Model H155. A "Current information" section notes that six additional clinical occurrences have been reported since the July 14, 2005 PPHN, and that no additional failures have been reported since June 17, 2005. It also mentions design changes approved by the FDA to reduce the likelihood of failure in newly manufactured devices.

Safety in the use of CEIDS

Table I.
Proportion of Explanted Devices Stratified Per Center Volume. The Center has been Classified on the Basis of the Volume of Impacted Devices (>30; 10–30; <10 Units Per Center)

Impacted Devices Per Center	# Centers	Total Devices Impacted	Total Devices Explanted	Mean % Explanted Per Center	P Value
>30	19	1349	221 (16%)	16.7%	0.092
10–30	74	1269	278 (22%)	21.3%	
<10	154	567	98 (17%)	14.6%	



In absence of underestimation of the events, **a lower incidence than expected** could resize the dimension of the problem, **justifying the concept of a more frequent follow-up** of patient with respect to the choice of an immediate device explant


Safety in the use of CEIDS

- ✓ Medtronic **Sprint Fidelis** implantable cardioverter-defibrillator leads are **prone to fracture**, which caused **inappropriate shocks and a few reported deaths**. When compared with Medtronic Sprint Quattro leads, they showed a significantly higher fracture rate (2.81%/year vs 0.43%/year, $P, 0.0001$).

Table 2. Fidelis and Quattro Lead Status and Failure Rates

	Fidelis (n=1023)	Quattro (n=1668)
Active, n (%)	683 (66.8)*	1290 (77.3)
Failed, n (%)	80 (7.8)*	23 (1.4)
Removed from service, n (%)	240 (23.5)	323 (19.4)
Died	145 (14.2)†	245 (14.6)
Replaced/abandoned	61 (6.0)*	24 (1.4)
Transplanted	10 (1.0)	14 (0.8)
Infection	12 (1.2)	23 (1.3)
Deactivated	12 (1.2)	17 (1.0)
Lost or followed up elsewhere, n (%)	20 (2.0)	32 (1.9)
Average follow-up, y	2.78*	3.18
Implant time, y	2844	5309
Failure rate, %/y	2.81*	0.43

* $P < 0.0001$; † $P = 0.76$.



Safety in the use of CEIDS

- ✓ During the study, no deaths occurred as the result of lead failure.
- ✓ Overall, 42% of lead failures (42 of 103) were associated with inappropriate shocks, including 37 Fidelis pace-sense conductor fractures
- ✓ Interestingly, **the failure rate was higher** for **younger patients**, for **women**, and for patients affected by **channelopathies** or by **hypertrophic cardiomyopathy**: these findings support once more the **need for risk stratification** in choosing whether to replace a lead or not

Safety in the use of CEIDS

	Sprint Fidelis		Quattro Secure	
	Failure Rate	95% CI	Failure Rate	95% CI
HCM	74.0	(36.9, 132.4)	3.2	(0.1, 17.7)
ARVD and channelopathies	68.7	(25.2, 149.4)	0.0	(0.0, 58.1)
Female	39.1	(25.8, 56.9)	2.9	(0.6, 8.4)
Ischemic HD	27.1	(19.6, 36.7)	4.4	(2.5, 7.2)
Idiopathic VT/VF	26.3	(3.2, 94.9)	0.0	(0.0, 14.2)
DCM	17.1	(9.3,28.6)	5.7	(2.1, 12.4)

HD indicates heart disease; VT, ventricular tachycardia; VF, ventricular fibrillation; and DCM, dilated cardiomyopathy.

Safety in the use of CEIDS

- ✓ In 1994,
- ✓ **Teletronics Accufix active fixation** leads were subjected to recall because of the risk of **fracture and protrusion of the J retention wire**, which had caused **pericardial tamponade, perforation of the right atrium, embolization** to the pulmonary circulation, and **a few deaths**.
- ✓ **Of the potentially affected leads, 13% were extracted, but the risk of fatal and of life-threatening complications were much higher for extraction than for leaving the lead in place**, especially for elderly patients, while the risk of lead fracture and complications were lower in the elderly.

Safety in the use of CEIDS

TABLE 3. WWR Reported Accufix Injuries Related to J Retention Wire

Injury	n
Death	6
Pericardial tamponade (nonfatal)	19
Pericardial effusion without tamponade	5
Atrial perforation with pericarditis	3
Embolism of J-wire fragment	4
Tricuspid valve perforation and insufficiency	1
Aorta–right atrial fistula	1
Right atrial thrombus with pulmonary thromboembolism	1

Total injuries=40.

Conclusion; Low, ongoing risk of injury.
Extraction associated risk is higher than conservative management approach.



Efficacy and safety in the use of CEIDS

- ✓ In fact, for both leads and devices, **the decision to replace an element subject to recall is still uncertain**, as it depends on several variables, among which the most important are the estimated rate of device failure, the arrhythmic risk, and the mortality rate of device replacement.
- ✓ In most cases, **a conservative strategy with short (3-month) follow-up may prove to be safer than an attempt to replace the potentially affected element**, either because the effective failure rate can be overestimated, or because of the low benefit/risk ratio for the replacement strategy.


Patient-focused perspectives such as quality of life, patient satisfaction/acceptability

- ✓ Clinicians are usually unaware of the **psychosocial impact of implanted PMs and ICDs**, as they focus their attention mainly on the technical aspects of device functions rather than on psychosocial factors.
- ✓ For this reason, a better understanding of factors likely to contribute to patients' perception of their own health would be help in their management, and in the training of clinicians and nurses.

Patient-focused perspectives such as quality of life, patient satisfaction/acceptability

- ✓ One of the first studies to assess differences in psychosocial adaptation, QOL, and incidence of affective disorders between patients with PMs and those with ICDs (shock and no-shock groups) found no difference in scores such as the Hospital Anxiety and Depression (HAD) scale or Short Form 36 QOL measurement.
- ✓ However, **ICD patients who had experienced a shock** reported more **limitation in their leisure time activities** than patients in the other two groups (ICD without shock and PM) and were particularly **concerned about the battery running out** or about **possible technical failures of their device**. A greater demand for a support team in the shocked ICD group than in the non-shocked ICD group or the PM group was also reported.

Ethical aspects of deactivating implanted cardiac devices



Endless shock, Endless pacing

Patient-focused perspectives such as quality of life, patient satisfaction/acceptability

- ✓ **Deactivation of ICDs** in patient nearing end of life is a controversial and debated issue, which is becoming increasingly topical. In this regard, the **Heart Rhythm Society, in its 2008 consensus statement on CIEDs**, affirmed that “The primary aim behind the rationale for deactivation must always be to **respect the patient’s right to live, or at least to die with dignity**, while limiting any therapeutic action that increases the patient’s level of stress, pain or anxiety.”
- ✓ More recently, the **European Heart Rhythm Society** issued a consensus statement on the subject of ICDs deactivation in patients with irreversible or terminal illness. In this document, the members of the EHRA committee have discussed the ethical, legal and technical aspects of this issue, following the “**key principles of liberal democratic societies, which include respect for the diversity of values and cultures, equal rights for all individuals, and preservation of fundamental human rights.**”

Wilkoff BL, et al. *Europace*. 2008;10:707–725

Padeletti L, et al. *Europace*. 2010;12:1480–1489

50Y, F

- C/C

- Swollen face and neck

- P/I:

- 1982: VVI PM (1st) for SSS

- 1986: 2nd PM (Teletronics, OPTIMA MP) for lead fracture



- 1992: 3rd PM (Teletronics, TEMPO VR 1102, VVIR) for recurrent lead fracture

50Y, F


■ P/I

- 1999: 4th PM for power depletion
- 2000: Painful swelling of right clavicular area
(Osteomyelitis of right clavicle)
5th pacing lead through left subclavian vein
- 2003: Swollen face and neck
Thrombolysis with 1,500,000 unit urokinase
Heparin and warfarin
- Jan. 2004: Intense facial and arm swelling

Chest X-ray and CT scan




Cavogram and balloon venoplasty



Pressure tracing

– Before and after venoplasty –



The patient required venous stenting because of recurrent SVC syndrome.

62Y, F

- C/C

- Swollen face, neck and left arm


- P/I:

- 2003; DCMP


- 2005; CRT-P

- 2006; Swollen face, neck and left arm

Venogram of left subclavian vein



Venoplasty



Thrombosis or venous stenosis

- ✓ Lead removal is recommended in patients with clinically **significant thromboembolism events** associated with thrombus on a lead or a lead fragment
- ✓ Lead removal is recommended in patients **with bilateral subclavian vein or superior vena cava occlusion precluding implantation of a needed transvenous lead**
- ✓ Lead removal is recommended in patients with **planned stent deployment in a vein already containing a transvenous lead**, to avoid entrapment of the lead
- ✓ Lead removal is recommended in patients with **superior vena cava stenosis or occlusion with limiting symptoms**
- ✓ Lead removal is recommended in patients with ipsilateral venous occlusion preventing access to the venous circulation for required placement of an additional lead when there is a contraindication for using the contralateral side (e.g., contralateral AV fistula, shunt or vascular access port, mastectomy)

Body Guard Saves Lives !!!



ICD saves lives !!!



ICD saves lives !!!

ICD User Group

Support, education and advocacy for recipients of Implantable cardioverter defibrillators.

myheart
mydata

[Home](#) [Glossary](#) [Mission](#) [Media](#) [Contact](#)

Friday, June 12, 2009

WATCH: ICD saves life of Belgium soccer player.

I don't know how much explaining is necessary here. This amazing video says it all.

In the clip, Anthony Van Loo, a 20-year old Belgium soccer player collapses during a match and is resuscitated by his Implantable Cardioverter Defibrillator (ICD). The device delivers a shock to restore his heart rhythm.



Most of the press has been reporting the incident as a [heart attack](#). This is not accurate. Instead, Van Loo must have suffered what is called an "arrhythmia". Arrhythmias are disturbances in the normal heart rate and electrical rhythm, and are usually life-threatening. Two of the most dangerous types of arrhythmia are called Ventricular Tachycardia (VT) and Ventricular Fibrillation (VF). Such abnormally fast heart rhythms prevent the heart from pumping blood to the brain, resulting in loss of consciousness. If left untreated, these dangerous rhythms will deteriorate into a cardiac arrest. [Read about the difference between cardiac arrest and heart attack.](#)

According to this [Time story](#) ([Saving Athletes from Cardiac Arrest](#), by Carolyn Sayre), "Sudden Cardiac Arrest [...] affects more than 400,000 people in the U.S. and is the leading cause of death in competitive athletes."

Search:

Search

Find us on
Facebook

Watch us on YouTube

Hugo Campos
HugoOC

HugoOC This one is for @MedtronicCEO; youtube.com/watch?v=3J0x5Z... #ShowMeTheData #s4pm yesterday · reply · retweet · favorite

HugoOC Medtronic tops the list of #medtech lobbying spenders with \$4.9M in campaign contributions for 2012 #meddevice ow.3y/38AQ 2 days ago · reply · retweet · favorite


dannunzio Data War Reaches The Supreme Court - Forbes onforb.es/ZvWAp CC: @phillipdave @HugoOC #bcm #hcam #digitalhealth 5 days ago · reply · retweet · favorite

HugoOC .@MedtronicCEO:




ICDs in the reduction of SCD in primary and secondary prevention

MADIT II




No. AT Risk					
Defibrillator	742	503 (0.91)	274 (0.84)	110 (0.78)	9
Conventional	490	329 (0.90)	170 (0.78)	65 (0.69)	3


SCD-HeFT



CIDS



AVID



CRT-D and CRT-P can reduce overall and CV mortality and number of hospitalizations

Comparison of Medical Therapy, Pacing, and Defibrillation in Heart Failure (COMPANION)






Table 2. Changes in Distance Walked in Six Minutes, Quality of Life, and New York Heart Association (NYHA) Class.

Variable	3 Months			6 Months		
	No. of Patients	Mean ±SD Change	P Value*	No. of Patients	Mean ±SD Change	P Value*
Increase in distance walked in 6 min (m)						
Optimal pharmacologic therapy	170	9±84	—	142	1±93	—
Cardiac-resynchronization therapy						
Pacemaker	422	33±99	<0.001	373	40±96	<0.001
Pacemaker-defibrillator	420	44±109	<0.001	378	46±98	<0.001
Increase in quality of life (%)†						
Optimal pharmacologic therapy	243	-9±21	—	207	-12±23	—
Cardiac-resynchronization therapy						
Pacemaker	510	-24±27	<0.001	460	-25±26	<0.001
Pacemaker-defibrillator	514	-24±28	<0.001	478	-26±28	<0.001
Improvement in NYHA class symptoms (%)						
Optimal pharmacologic therapy	242	24	—	199	38	—
Cardiac-resynchronization therapy						
Pacemaker	551	58	<0.001	489	61	<0.001
Pacemaker-defibrillator	543	55	<0.001	497	57	<0.001

CARE-HF study




Bristow MR, et al. *N Engl J Med.* 2004;350:2140–2150



Cleland JGF, et al. *N Engl J Med.* 2005;352:1539–1549

Cost-effectiveness of CEID – MADIT II

- ✓ Advantage of the use of ICDs in preventing SCD increases further at **long-term follow-up (8 years)**, while the cost-effectiveness ratio per saved discontinued life-year dramatically improves during the same period





“... 무서워 잠 못담글까?”

**CIED implantation:
Is it a just beginning of problems?**

Yes!!!

**Just the beginning of lots of events
which we need to overcome!!!**